

CERTIFICATE

No.: TM 61000575.002



HOGER

Licence holder

HogerBox Sp. z o. o.
ul. Lipowa 3
30-702 Kraków

Manufacturing Plant

HogerBox Sp. z o. o.
ul. Lipowa 3
30-702 Kraków

Project number

26100611

Our reference

SD/84953490

Expiration date

from 03.12.2021 to 02.12.2026

Basis of research

PN-EN 12469: 2002

TÜV Rheinland Polska Sp. z o.o. declares that the product described below meets the requirements contained in the reference documents:

Microbiological safety cabinet (MSC) class III

HOGER HOG/S/16/OP

Technical parameters of the glovebox:

- performance class for cleanability: CI-B,
- performance class for sterilizability: SI-B,
- working pressure: < -200 Pa,
- air exchange rate: > 3 volumes/min,
- airflow through open gloveport: > 0.7 m/s,
- filtration efficiency: > 99.995%.

TÜV Rheinland Polska Sp. z o.o.

ul. Wolności 347,
41-800 Zabrze, Polska
Tel.: (+48/32) 271 64 89
e-mail: post@pl.tuv.com



Product certification body


Tomasz Opaszowski

Zabrze, 04.03.2022

This certificate is subject to the Certification Terms and Conditions and the JCW TRP General Transaction Conditions and applies only to the products that are compliant with the standard used for compliance assessment. This certificate alone does not entitle the holder to affix the CE mark.
This certificate entitles the holder to affix the product with the TÜV mark



Safety
Regular
Production
Surveillance



www.tuv.com
ID 000081619



TÜVRheinland®
Precisely Right.

www.tuv.pl